



Nederlands Meetinstituut

Test certificate

Number **TC7515** revision 0

Project number 810183

Page 1 of 4

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body Number 0122

In accordance with Paragraph 8.1 of the European Standard on Metrological aspects of non-automatic weighing instruments EN 45501:1992/AC:1993 and by application of the OIML International Recommendation R 60 (Edition 2000). The applied error fraction p_j , meant in the paragraph 3.5.4. of the standard is 0.7.

Applicant BAYKON Endustriyel Kontrol Sistemleri A.S.
Tuzla Organize Sanayi Sitesi
Organik Sokak 31, Tepeoren, Tuzla
34956 Istanbul
Turkey

In respect of A **shear beam load cell**, with strain gauges, tested as a part of a weighing instrument.

Manufacturer : BAYKON Endustriyel Kontrol Sistemleri A.S.
Type : BS555

Characteristics E_{max} : 50 kg up to and including 250 kg
Accuracy class : C3

In the description number TC7515 revision 0 further characteristics are described.

Description and documentation The load cell is described in the description number TC7515 revision 0 and documented in the documentation folder TC7515-1, appertaining to this test certificate.

Remarks Summary of the test involved: see Appendix number TC7515 revision 0

Dordrecht, 19 December 2008
NMI Certin B.V.


Ing. C. Oosterman
Manager Product Certification

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (27228703)
NMI Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI, B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission.