



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.16
Project number 2173264
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: C. Oosterman



**Applicant and
Manufacturer**

Quantronix Inc.
380 South 200 West
P.O. Box 929
84025 Farmington, Utah
United States of America

Identification of the
certified type

A Non-automatic weighing instrument

Type : Cubiscan 100-LFT
Cubiscan 100-TLFT

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



OIML R 76 - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
15 March 2019



C. Oosterman
Head Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org





OIML Member State
The Netherlands

OIML Certificate

Number R76/2006-A-NL1-19.16
Project number 2173264
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-14200522-01 dated 3 December 2014 that includes 55 pages;
- No. NMI-2173264-01 dated 13 March 2019 that includes 56 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max \leq 50 kg
Verification scale interval	e \geq 20 g
Weighing range	Single interval
Maximum number of scale intervals	n \leq 2500
Temperature range	-10 °C / +40 °C
Power supply voltage	100 – 240 V AC 50/60 Hz
Software identification	Version number 4.xxx (x = 0... 9) or 5.xxx (x = 0... 9)