



Nederlands Meetinstituut

# Test certificate

Number TC2097 revision 4  
Project number 10092781  
Page 1 of 4

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Notified Body Number 122

In accordance with Paragraph 8.1 of the European Standard on Metrological aspects of non-automatic weighing instruments EN 45501:1992/AC:1993 and by application of the OIML International Recommendation R 60 (Edition 1991). The applied error fraction  $p_i$ , meant in the paragraph 3.5.4. of the standard is 0.7.

Applicant Flintec GmbH  
Bahnhofstraße 52-54  
74909 Meckesheim  
Germany

In respect of The model of a compression load cell, with strain gauges, tested as a part of a weighing instrument.  
Manufacturer : Flintec or Flintec  
Type : RC1

## Characteristics

Maximum capacity ( $E_{max}$ )	250, 400, 600 and 900 kN	
Accuracy Class	C	
Maximum number of load cell intervals (n)	1000	3000
Ratio of minimum LC Verification Interval $Y = E_{max} / V_{min}$	4667	10000

In the description TC2097 revision 4 further characteristics are described.

Nederlands Meetinstituut  
Hugo de Grootplein 1  
3314 EG Dordrecht  
Telephone +31 78 6332332  
Telefax +31 78 6332309

NMI B.V. (Chamber of Commerce Haaglanden  
No. 27228701)

Subsidiary companies:  
NMI Certin B.V. (27233418)  
NMI Van Swinden Laboratorium B.V. (27228703)  
NMI International B.V. (27239176)

This document is issued under the provision that NMI B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission.



QUALIFIED  
BY STERLAB  
Reg. nr. L 029